

CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361

Lot No.: 20240502Y

Expiry date: Aug. 01, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12700	≥9100, ≤16900
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	13400	≥9100, ≤16900

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.



 Ryosuke Sugimoto; Manager

Date May. 27, 2024

KBG6036100ac-Eaa
 SA2405-0297

Quality Assurance Group, Edogawa
 Quality Unit
 Kikkoman Biochemifa Company