



CERTIFICATE OF ANALYSIS

LuciPac A3 Water

Product code :60365

Lot No.: 20240909W

Expiry date: Dec. 08, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12700	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	13500	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H Katayama
Hiroshi Katayama; Manager

Date Oct. 10, 2024

KBG6036500ac-Eaa
SA2410-0175

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company