kikkoman Kikkoman Biochemifa Company

CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361

Lot No.: 20260729Y Expiry date: Oct. 28, 2027

TEST PARAMETER	RESULT	SPECIFICATION
When applied 100µL of 5pmol ADP Average Luminescence(RLU)	12500	≧9100, ≦16900
When applied 100µL of 5pmol AMP Average Luminescence(RLU)	12600	≧9100, ≦16900

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H. Kikuchi

Hisako Kikuchi; Manager

Date Aug. 24, 2026

KBG6036100ac-Eaa
SA2608-0338

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company